

# Analysis of Greenhouse Gasses by GC/MS

Emily M. Bock – Virginia Tech

Zachary M. Easton, Ph.D. – Virginia Tech

Paul V. Macek – Shimadzu Scientific Instruments, Inc.


This work was a cooperative effort between Shimadzu Scientific Instruments and the Biological Systems Engineering Department at Virginia Tech.




# Objective

- Develop a method that uses common laboratory instrumentation
- Existing instrumentation is dedicated to GHG analysis only
- Dedicated GHG analyzers are complex, multicolumn instruments
- Most modern labs are equipped with a GC/MS

# GHG Analyzer



# GC/MS QP-2010 Ultra at Virginia Tech





# GC Parameters

- Injection Port Temperature: 200 °C
- GC Oven Program
  - Initial Temp: 100°C
  - Initial Hold: 0 min
  - Program Rate 35°C/min
  - Final Temp: 180°C
  - Post Run Temp: 240°C for 1 min
  - Carrier Gas: Helium
  - Linear Velocity: 45 cm/sec
  - Flow Mode: Constant Linear Velocity
  - Split Ratio: 5
  - Injection Volume: 1 mL
  - Column: 60m Carboxen® 1010 PLOT (Supelco) 0.32 mm id


# MS Parameters

- MS Interface Temperature 220°C
- MS Source Temperature 200°C
- Electron Multiplier Voltage:
  - CH<sub>4</sub>: 1.25 kV
  - CO<sub>2</sub>: 0.85 kV
  - N<sub>2</sub>O: 1.25 kV
- Scan Parameters
  - CH<sub>4</sub>: SIM for m/z 16 and m/z 15; Quant from m/z 16
  - CO<sub>2</sub>: Scan from 15 to 50 AMU; Quant from m/z 44
  - N<sub>2</sub>O: SIM for m/z 30 and m/z 44; Quant from m/z 30

# CH<sub>4</sub> Calibration Curve

$$R^2 = 0.9999$$

$$R = 0.9999$$






# CO<sub>2</sub> Calibration Curve

$$R^2 = 0.9999$$


$$R = 0.9999$$




# N<sub>2</sub>O Calibration Curve

$$R^2 = 0.9996$$


$$R = 0.9998$$



# GC/MS Chromatogram of GHG Standard



# GC/MS Chromatogram of Outside Air



# Method Detection Limits

Calculated per 40 CFR Part 136 Appendix B

Methane MDL = 0.05 ppmv

Nitrous Oxide MDL = 0.02 ppmv

# Environmental Samples

## Ambient Air Results Sample 1

Methane	1.6	ppmv
Carbon Dioxide	397	ppmv
Nitrous Oxide	0.27	ppmv

## Ambient Air Results Sample 2

Methane	1.6	ppmv
Carbon Dioxide	394	ppmv
Nitrous Oxide	0.30	ppmv

Air samples were taken on the campus of Virginia Tech in the Ag Quad



# Acknowledgments

- 40 CFR Part 136 Appendix B
- Partial support provided by the Virginia Water Resources Research Center and the Virginia Tech Institute for Critical Technology and Applied Science
- Thanks to Jeffery Parish of Shimadzu Scientific Instruments, Inc. for providing a viable manual injection technique.
- Thanks to the Luxbacher Lab at Virginia Tech Mining and Minerals Engineering for assistance with developing gas sampling techniques.