

IMPROVING CONFIDENCE IN THE TO-15 ATMOSPHERIC ANALYSIS AT TRACE CONCENTRATIONS

Whipple, Wayne J., Michelle Kerr, Troy Strock,

US EPA R5 CRL

Chris Harmon

Humboldt State University, Chemistry Dept.

Disclaimer

The information presented here does not necessarily represent the views of the United States Environmental Protection Agency.

There is no endorsement of any products mentioned or shown in this presentation. They are presented for methodology information only.



Why is this talk different than a TO-15 Method Talk and why do we need trace levels?

- Many TO-15 analysis have high reporting limits.
 - 100 pptv and higher.
 - Risk based target concentrations can be extremely low << 50 pptv.
- Modeling typically uses $\frac{1}{2}$ the detection level as a value overestimating many compounds.
- National Air Toxics Trends Study (NATTS) is targeting lower concentrations in their program.
- Achieve ambient air concentrations.

Indoor Vapors Intrusion Calculator

	µg/m ³	pptv
1,1,2,2-tetrachlorethane	0.048	7
tetrachloroethene	1.11	162
trichloroethene	0.48	89
1,1,2-trichlorothane	0.18	33
vinyl chloride	0.17	67
1,2-dibromoethane	0.0047	1
benzyl chloride	0.057	11
benzene	0.36	100
acrolein	0.021	9

*There are no uncertainties listed for these target values. These values are personal communication with risk assessors and are orders of magnitude lower than the values used in the risk assessments.


Ambient Air Health Comparison Values

Compound	CAS	Risk-based health comparison values for chronic exposures (lowest value shaded orange)		
		1-in-1-million cancer risk, ppt	10-in-1-million cancer risk, ppt	Noncancer effects, ppb
benzene	107-02-8	n/a	n/a	0.009
chloroethane; 45850	91-20-3	5.6	56	0.57
ethylene dichloride	78-87-5	11	114	0.87
isobutadiene	106-99-0	15	151	0.90
methacrylonitrile	107-13-1	6.8	68	0.92
vinylene dibromide	106-93-4	0.22	2.2	1.2
trans-1,2-dibromoethyl bromide	74-83-9	n/a	n/a	1.3
ethylene	71-43-2	39	393	9.2
hexachloroethane	56-23-5	26	265	16
chloroform	67-66-3	n/a	n/a	20
hexanes	1330-20-7	n/a	n/a	23
methane chloride	75-01-4	44	437	38
trichlorobenzene	106-46-7	15	152	133
toluene	110-54-3	n/a	n/a	198
benzene	100-41-4	92	922	230
chloroform	71-55-6	n/a	n/a	926
methane	108-88-3	n/a	n/a	1326
1,2-tetrachloroethane	79-34-5	2.5	25	n/a
ethylene dichloride	107-06-2	9.5	95	n/a
trichloroethylene	127-18-4	25	250	n/a
tetrachloroethylene	79-01-6	93	931	n/a
vinylene chloride	75-09-2	608	6079	n/a

Concentrations are derived from the cancer Unit Risk Estimates (UREs) from IRIS and other sources.

Achieving these limits with high confidence

- The achieve these low limits most laboratories and project planners will have to balance added **confidence in the results** against the **expense and time** of the rigorous procedures to obtain the confidence.



\$\$\$\$

\$\$\$

What are the limiting factors for ultra trace concentrations?

Instrument limitations (single quadrupole full scan or selective ion monitoring (SIM), Time of Flight (TOF))

Background in makeup and carrier gasses, He, N₂, Zero Air

Cleanliness and integrity in:

- regulators and standard makeup (including internal standards and surrogates if added)
- sample preparation, preconcentrator
- canisters or alternative containers
- sample train (passive regulators, connections, transfer lines, pumps)

Accuracy of calibration procedures

Carry over in instrument

- Naphthalene and other compounds with high vapor pressures

bid SIM unless very confident from interference

Often 1,3-butadiene may be mostly n-butane secondary ion, no unique ion between them

Acrolein can be product of a butene secondary ions.

Other interferences possible and can only be sure in full scan mode

Interferences are not in standards or performance testing samples.

What solutions are used to engineer out, minimize or deal with contamination interferences?

dedicated and clean regulators for standards

- internal standard, 1 ppmv standard, varying analytes

more rigorous canister cleaning procedure

certify each canister and not by batch

- selective batch certification as next best approach
- match certification selection results to sample results

calibrate using a linear or quadratic force through zero curve fit when possible

and a % error acceptance criteria at each calibration level

check integrity of canisters with low concentration spikes

- not routinely performed at CRL


If an environmental chamber is available, use zero air to test regulators and

sample train, right now CRL uses lab air

Peggy

cold trap dehydration

Amelia



pegasus HT, Agilent 7890,
7100, Entech 7016
compounds 1 to 5 pg > 5 S/N FS



LECO TruTOF, Agilent 7890,
Entech 7200, Entech 7650
Most compounds 1 to 5 pg > 5

Chamber Cleaning

° C

vacuum to 50 mtorr

flush to 20 psi whole air

2 minutes

repeat 10- 20 times

90 °C for 3 days open to room air
flush immediately with zero air



Canister Storage

Canisters in zero air 20 psia
Quantitation report of blank cert.,
Report with data file, and match to sample,
detected in blank and canister > 5 times blank



	MAX, pptv	n	Average, pptv
ne	232	13	36
	84	17	30
	83	6	34
chlorobenzene	54	13	25
robenzene	37	15	4
yl ketone	19	16	7
I	18	17	6
oride	17	1	
	16	10	13
e	15	15	0
	12	17	-1
le Chloride	12	17	5
robenzene	11	17	2
e	9	6	4
	9	16	4
obutadiene	9	17	3
	9	17	2
ethylbenzene	8	11	2
robenzene	8	10	2
fluoromethane	7	11	2

Canister Blank
 3 day oven clear
 Other compounds
 under a max
 concentration of
 5 pptv.
 17 cans some from sc

	MAX pptv	average, pptv	count
aceton	351	151	21
ol	295	242	21
lene	250	79	21
lene	109	-27	20
2-Dichloroethene	53	14	12
chlorobenzene	38	10	21
acetate	30		15
	30	20	20
ethyl ketone	29	17	21
oroethene	27	3	21
	23	13	21
difluoromethane	22	12	18
	21	4	21
chloride	15	5	13
oroethane	14		4
robenzene	12	4	21
one	12	10	8
ethene	11	4	17
3-Dichloropropene	10	7	12
robutadiene	9	-1	21
robenzene	9	2	21
romoethane	8	3	17
rofuran	7		2
ichloropropene	7	5	9
robenzene	7	1	19
	7	-6	21

Canister Blank Data
 20 x cycle cleaning
 Other compounds
 under a max
 concentration of
 5 pptv.
 21 cans

Canister Integrity?

One lab rigorously cleaned new canisters delivered to lab and tested each with zero air.

- Compounds still present in one canister, vinyl chloride, benzene and acrolein at 100 pptv or less

Filled canisters with 500 pptv and elevated acrolein and depressed 1,3-butadiene and dichloropropenes.

How to clean, and
calibrate sampling
systems?
The quality of sample is
only as good as the
weakest link in the
sample path

from Entech Instrument Catalogue

RS INC.



E Passive Canister Sampler

inert, compact, and reliable flow path with superior low flow

CS1200ES (Silonite® coated)

Calibration Curve Fits

use linear or quadratic

force through zero whenever possible

Makes it easy to see the interferences are present

Greatly improves accuracy at low concentrations


eight calibration curve when not forcing through zero

check with % error between expected and calculated from calibration curve


$\pm 50\%$ error at RL, $\pm 30\%$ all points above,

Most levels should be tighter to about $\pm 10\%$

2 ppbv calibration standard

Rtx-VMS 30 m 0.25 mm ID, 1.4 μ m

200 cc Integrated air sample



1,1,1-trichloroethane

Standard

150203cal1 2 pptv:2

150203cal2 5 pptv:2

150203cal3 10 pptv:2

150203cal4 20 pptv:1

150203cal5 50 pptv:1

150203cal6 100 pptv:1

150203cal7 250 pptv:1

150203cal8 500 pptv:1

150203cal9 1000 pptv:1

150203cal10 2000 pptv:1

Area

56424

143372

[260115](#)

456658

1074617

2144115

5614373

11203275

22478983

44979119

Quant S/N

56

134

242

442

1019

2040

5216

10475

20642

40030

Cert. Conc.

2.0

5.0

10.0

20.0

50.0

100.0

250.0

500.0

1000.0

2000.0

Calc. Conc.

2.2

5.8

10.7

19.0

45.6

91.3

240.5

485.3

989.6

2010.6

% Diff. Conc.

12.0

16.0

7.3

5.0

8.8

8.7

3.8

2.9

1.0

0.5

Weighting

1

1

1

1

1

1

1

1

1

1

RL

150203cal10 2000 ppt

150203cal9 1000 pptv:1

150203cal8 500 pptv:1

150203cal7 250 pptv:1

150203cal1 2 pptv:2

200

400

600

800

1000

1200

1400

1600

1800

Concentration


$$y = +0.00271438x + 0$$


$$r = 0.99994$$

Forced zero intercept

Carbon disulfide



Standard	Area	Quant	S/N	Cert.	Conc.	Calc.	Conc.	% Diff.	Conc.	Weighting	
150203cal1 2 pptv:2	7549067	5868.4		2	215.18		10659		1		150203cal10 2000 pptv:1
150203cal2 5 pptv:2	6806342	5181.6		5	188.75		3675		1		
150203cal3 10 pptv:2	6567250	5226		10	183.69		1736.9		1		
150203cal4 20 pptv:1	6508335	5108.8		20	178.39		791.95		1		
150203cal5 50 pptv:1	7472549	5796.8		50	204.82		309.64		1		
150203cal6 100 pptv:1	9118154	7168.1		100	266.82		166.82		1		
150203cal7 250 pptv:1	13934292	10969		250	404.11		61.643		1		
150203cal8 500 pptv:1	21693020	17886		500	630.16		26.032		1		
150203cal9 1000 pptv:1	36450771	28878		1000	1052.3		5.228		1		RL
150203cal10 2000 pptv:1	65844432	46714		2000	1906.7		4.664		1		






Box and Whisker Plot Legend For blank canister data

opene Calibration Curves w and w/o forced zero



Blank Canister Contamination in pptv

n=21



Propene calibration


Standard	Area	Quant S/N	Cert. Conc.	% Diff.			% Diff.		
				Calc. Conc.	Conc.	Weighting	Calc. Conc.	Conc.	
150203cal1 2 pptv:2		168794	54.629	2	22.1	1005	0.5	3	65
150203cal2 5 pptv:2		169726	59.511	5	21.6	332	0.2	3	44
150203cal3 10 pptv:2		229931	76.979	10	29.5	195	0.1	11	11
150203cal4 20 pptv:1		287833	100.74	20	36.2	81	0.05	18	9
150203cal5 50 pptv:1		467081	154.08	50	58.8	18	0.02	42	16
150203cal6 100 pptv:1		842995	266.54	100	113.3	13	0.01	99	1
150203cal7 250 pptv:1		1880888	590.75	250	250.5	0	0.004	243	3
150203cal8 500 pptv:1		3583291	1062	500	478.1	4	0.002	482	4
150203cal9 1000 pptv:1		7245002	1983	1000	960.6	4	0.001	989	1
150203cal10 2000 pptv:1		14783116	3743.9	2000	1966.2	2	0.0005	2046	2

Force Zero

RL

Calculated Intercept RL


chloromethane global temporal data and a sample with two calibration methods




cal name	Sample Conc
141028 CFC lin 0 1W	
141028 CFC linear	
141028 CFC lin 0 1W	
141028 CFC linear	
141028 CFC lin 0 1W	
141028 CFC linear	
141028 CFC lin 0 1W	
141028 CFC linear	
141028 CFC lin 0 1W	
141028 CFC linear	
141028 CFC lin 0 1W	
141028 CFC linear	
141028 CFC lin 0 1W	
141028 CFC linear	

Data from one week at site i
Calibration based RL 50 pptv
 $\langle \text{blank} \rangle_{\text{0int}} = 18 \text{ pptv}, n=20$
 $\langle \text{blanks} \rangle_{\text{linear}} = 0.6 \text{ pptv } n=2$

Two compounds that are in every ambient air sample at very low concentrations





mL air sample




7 p

Freon 12 (dichlorodifluoromethane) vs. Year






locarbon lobal Trends




2002 Chicago Study for Freons as Surrogates




Chicago Halocarbon Concentrations, October 2004 through April 2005





Freon Data for Indiana Project n=40



Sample Monitoring Compounds (Freons)

Data Over 2 Month Sampling Mission

pptv	1,1,1-TCA	F-11	CCl4	F-12	F-113	F-114
NOAA-HATS	4	235	83	510	72	
Average ~40 samples	4.7	227.4	82.5	469.1	73.9	20.6
% accuracy	18%	-3%	-1%	-8%	3%	
% RSD	27%	9%	13%	11%	23%	18%

Results of Freons with Various Curve Fits

	Quantification pptv	Sample	expected	% difference	pptv	Sample	expected	% difference
Average RF	1,1,1-Trichloroethane	3.33	4	17%	Carbon Tetrachloride	61.7	83	
lin 0 1W	1,1,1-Trichloroethane	4.05	4	-1%	Carbon Tetrachloride	71.1	83	
lin 0 invW	1,1,1-Trichloroethane	4.01	4	0%	Carbon Tetrachloride	70.2	83	
linear	1,1,1-Trichloroethane	-2.62	4	166%	Carbon Tetrachloride	62.6	83	
linear invW	1,1,1-Trichloroethane	-1.81	4	145%	Carbon Tetrachloride	64.6	83	
Average RF	1,1,2-Trichloro-1,2,2-Trifluoroethane	61.3	72	15%	Dichlorodifluoromethane	348	510	
lin 0 1W	1,1,2-Trichloro-1,2,2-Trifluoroethane	74.6	72	-4%	Dichlorodifluoromethane	423	510	
lin 0 invW	1,1,2-Trichloro-1,2,2-Trifluoroethane	72.5	72	-1%	Dichlorodifluoromethane	407	510	
linear	1,1,2-Trichloro-1,2,2-Trifluoroethane	53.6	72	26%	Dichlorodifluoromethane	401	510	
linear invW	1,1,2-Trichloro-1,2,2-Trifluoroethane	61.6	72	14%	Dichlorodifluoromethane	405	510	
Average RF	Dichlorotetrafluoroethane	19.7	20	2%	Trichlorofluoromethane	171.0	235	
lin 0 1W	Dichlorotetrafluoroethane	19.7	20	2%	Trichlorofluoromethane	192.4	235	
lin 0 invW	Dichlorotetrafluoroethane	19.8	20	1%	Trichlorofluoromethane	190.6	235	
linear	Dichlorotetrafluoroethane	19.7	20	2%	Trichlorofluoromethane	185.5	235	
linear invW	Dichlorotetrafluoroethane	19.8	20	1%	Trichlorofluoromethane	187.2	235	

response factor, lin 0 1W = linear forced 0 equally weighted, lin 0 invW = linear forced zero intercept 1/x weighted, linear = linear 1/x weighted calculated intercept, linear invW = linear calculated intercept 1/x weighted

ng calibration force through zero vs. non force through

Force through zero

- Gives better results with very low concentration and extremely little background or inconsistent background.
- Provides ability to understand background interferences.
- May give false positives when not used properly and careful attention to contamination sources are not observed.

Calculated intercept

- Best used when a much more constant background is present.
- May supply false negatives and negative concentrations.
- Always best to engineer out interferences whenever possible.

What can we learn about the analysis using Freons as sample monitoring compounds?

- Sensitivity of the system is stable
- Calibration standards are made correctly (not necessarily fractionation issues with low vapor pressure compounds)
- Sample was collected from outside air and sampled properly
 - Be careful of starting and stopping pressure in canisters to ensure proper delivery of sample (room air also has same Freon concentration)
- If system is behaving linearly, improve calibrations for instrument

Summary

Trace Concentrations in air analysis costs money, time

Manage Data Quality Objectives to as high an RL as possible for project

Control interferences and contamination.

Carefully select calibration model for specific analytes depending on system conditions

Monitor blanks in canisters and system

Use banned CFC concentrations and other compounds in ambient samples to understand how system works and get idea of quality of sample

Appreciation

US EPA R5 Laboratory personnel for support

Dr. Daniel Riemer for data and information

US EPA R5 Air and Radiation Division especially Dr. Motria Caudill